

CSK Himachal Pradesh Krishi Vishvavidyalaya, Palampur
Central Purchase & Disposal Unit
Office of the Comptroller

No. QSD.CPDU/4-20/CSKHPKV/-

13946-56

Dated:

Notice Inviting Quotations for the purchase of Ultra Centrifuge

Sealed Quotations are invited by the Comptroller, CSK HPKV, Palampur for the purchase of listed below equipment:

S.No.	Name of Item	Specifications	Quantity	Rate (in Rs.)
.1.	Ultra Centrifuge	Attached as Annexure-I	1 No.	

Term and Conditions

1. The quotation should be reach in the O/O the Comptroller, CSK HPKV, Palampur in sealed cover and super-scribed "**quotation for purchase of Ultra Centrifuge**" on 30-03-2021 by 5.00 PM. The quotation will be opened on 31-03-2021 at 11.00 AM in the presence of quote firms or their authorized representative, if present.
2. The rates should be quoted F.O.R. CSK HPKV, Palampur inclusive of all charges. GST extra, if any. (Nothing will be paid extra).
3. Price list should be attached.
4. The quantity of materials can be increased or decreased.
5. The Comptroller also reserves the rights to consider or reject any of the quotations without assigning any reason.
6. The supply of the material is to be made by the supplier immediately.
7. The quotation received after due date will not be considered.
8. The condition of advance payment or through cash etc. will not be considered.
9. The quotation should be submitted only if the items can be supplied promptly on demand.
10. The price quoted should be strictly according to the specifications stated above.
11. The payment will be made after satisfactorily Inspection Report of the Inspection Committee so constituted for the purpose by the University.
12. Each quotation should be accompanied with a refundable earnest money @ 2%, if the total amount for the supply of the material exceeds Rs. 1.00 lakh in the shape of fixed deposit receipts/TDR/DD in favour of the Comptroller, CSK HPKV, Palampur. **Without EMD the quotations will not be considered.**
13. The supplier will adhere to the provision of Motor Vehicle Act for transportation of material.
14. Performance security will be obtained from successful bidder @ 5% of the awarded material.
15. GST may be shown separately, copy of GST registration be enclosed with quotation and GST No. may be mentioned in the quotation.
16. The E-way bill is to be required alongwith actual bill.
17. Authorization Dealer Certificate and warranty certificate duly attested must be attached with quotation.
18. The quotation should accompany the satisfactory user report/certificate from the Institutions to whom the said equipment earlier supplied by the firm.
19. The conditional quotation will not be entertained.
20. The firms/suppliers shall observe all the guidelines/SOPs for COVIDI-2019 prescribed by the H.P. State Govt/Central Govt. from time to time.
21. All disputes shall be within the jurisdiction of Palampur Court.

M/S _____

all
Asstt.Registrar(CPDU),
CSK HPKV, Palampur.

CC:

1. The Head, Deptt. of Genetics and Plant Breeding, COA, CSK HPKV, Palampur.
2. Incharge, UNS, CSK HPKV, Palampur for uploading the NIQ in University Website.

Ultra centrifuge

Specifications

Bench Top High-Capacity Refrigerated Centrifuge:

- Stable speed output (max speed 6000rpm)
- High-Capacity (Maximum Capacity 400ml)
- Rotors & reduction adaptors (Angle Rotor 16 x 15 ml) with reduction adaptors
- Precise temperature controls up to - 8°C
- Pre-cooling feature
- Fast spin feature
- Calibration window on lid for speed (rpm)
- Brushless Induction motor with variable frequency drive ensures gentle start
- Microprocessor based unit with alpha numeric LCD display
- Self-diagnosis of errors
- Parameters setting through easy-to-use bi-directional encoder
- Inverter fault detection with auto shutdown
- CFC free refrigeration system
- 24 programme memory
- 3 Accelerations, 4 Deceleration profiles
- Choice of rpm / rcf setting & display
- Operating panel with feather touch keys
- Digital countdown timer 0-99 minutes & continuous run
- Last set parameters recall (Useful for repetitive analysis)
- Automatic door opening
- Log of cumulative centrifuge run time
- Automatic Rotor identification
- Safety lid interlock to prevent lid opening during centrifugation
- Imbalance detection & centrifugation stop with display of error
- Motor overload protection
- Gas hinge to prevent falling of door
- Emergency lid lock release

Sachin

06/03/2021
S. A. HT
6/3/2021

Basudra'
Co PJ PS (PB)
Dept of GPB
CSK HPKV Palampur

HASE
Dy No 4648
Date 06/03/2021